




# FullRange A-Constant conversion table

Optical / Immersion US  
Contact US

	<b>SRK/T</b> A Constant	<b>Hoffer Q</b> pACD	<b>Holladay 1</b> sf	a0	<b>Haigis</b> a1	a2
 BunnyLens	118.5	5.2	1.42	0.978	0.4	0.1
	118.16	4.98	1.2	0.753	0.4	0.1
 SeeLens	118.6	5.26	1.48	1.044	0.4	0.1
	118.26	5.05	1.27	1.819	0.4	0.1
 VisTor MF Plate Haptic	117.7	4.86	1.02	0.448	0.4	0.1
	117.3	4.61	0.77	0.184	0.4	0.1

1 IOL constant was evaluated using [IOL master and the SRK/T formula, relations between constants](#)

2 IOL constant was evaluated using [US biometry and the SRK/T formula, relations between optical and us biometry](#)

\* It is recommended that surgeons personalize their IOL constant based on their surgical techniques and equipment, experience and post-operative results.